Tehran, Tehransar

FRU®

پترو فرهان گستر جنوب

FGJ
FARHAN GOSTAR JOONOB

Make Measurement Easier   Professional Precise Portable

CE   RoHS   CNAS   SCM   GRGTEST 广电计量   ilac-MRA   SMQ

Tri-angle Gloss Meter WG68 (20, 60, 85 Degree)

Introduction:

WG68 gloss meter is produced according to the national verification regulation of JJG696 standard, adopt long life light source, good repeatability, small volume, high precision and good stability. It is mainly used coatings, printing ink, paint, paint, wood, marble, granite, to change polishing brick, brick of pottery and porcelain, plastic, paper, metal and non-metallic materials to the surface of the mirror gloss measurement, more stable performance, more durable, more comfortable;

Measuring Area Application:

20° (10*10mm) for high gloss, more than 200Gu.

60° (9*15mm) for semi gloss, from 10-200Gu.

85° (5*38mm) for low gloss, less than 10Gu.

Advantages :

1. One button to complete triangle measurement: one operation can complete three angle measurement; Measuring light incident angle meet ISO 2813 three measuring light incident angle. Three incident angle of mirror gloss meter is 20 °, 60 °, 85 ° respectively;

2. Optional angle mode: choose measuring angle or angle combination according to user needs.

3. Manually entering standard samples gloss data;

4. Humanized design, good man-machine communication interface;

5. Three angle can store 50000 testing records, storage structure is: 100 prototypes, each prototype with 500 testing data;

6. Data browsing and viewing function;

7. The machine has automatic shutdown function, low power warning functions, full data storage hint functions;

8. USB communication interface to communicate with PC;

9. PC software can be extended more functions.

+982165565901

+982144584619

+989034119385

پتروفرهان گستر جنوب

FGJ-NDT.IR
DIGINDT.IR

Tehran, Tehransar

FRU®

پترو فرهان گستر جنوب

FGJ
FARHAN GOSTAR JOONOB

Make Measurement Easier

Professional Precise Portable

CE   RoHS   CNAS   SCM   GRGT EST 广电计量   ilac-MRA   SMQ

Data Sheet

| Model	WG68
Measurement Range	20°: 0-2000 60°: 0-1000 85°: 0-160
Application	paint, ink, stoving varnish, coating, wood products; Marble, granite, vitrified polished tile, pottery brick and porcelain; Plastic, printing, paper; Other metallic material.
Projection Angle	20° 60° 85°
Measuring Spot	20°: 10x10 60°: 9x15 85°: 5x38
Measuring Mode	Automatic/manual/single/ average and combination angle/Show maximum/minimum value
Standard Sample Setting	Manually setting/bring out stored data for setting/manually setting tolerance
Data Storage	50000 groups set standard sample in any stored data, user-define the tolerance range
Language	English
Output	Software to PC
Division Value	0.1Gu
Repeatability	0-100:0.2 > 100:0.2%
Reproducibility	0.2Gu
Indication Error	1.0Gu
International Standard 60°	ISO-2813/ASTM-C584/ASTM-D523/DIN-67530/ASTM-D2457/JND-A60/JND-P60/ISO7668
National Standard 60°	GB 9754/GB 9966/GB 8807 - GB/T 9754,GB/T 13891,GB/T 7706,GB/T 8807
Working Temperature	10℃ - 40℃
Storage Temperature	-10℃ - 70℃
Humidity Range	Relative humidity less than 85%, without condensation
Volume	159x49x72mm
Weight	500g
Standard Accessories	Power adapter/User manual Warranty card/standard board/5V battery/USB cable /GlossW software (CD)

+982165565901
+982144584619
+989034119385

پتروفرهان گستر جنوب
FGJ-NDT.IR
DIGINDT.IR

پترو فرهان گستر جنوب

FGJ
FARHAN GOSTAR JOONOB

FRU®

Make Measurement Easier

Professional Precise Portable

CE RoHS CNAS SCM GRGTEST 广电计量 ilac-MRA SMQ

GlossW Software

GlossW - Untitled

File(F) Seting(S) Operate(O) Help(H)

DOCUMENT NO.	P-505146				
COMPANY	SHENZHEN xxxx CO.,LTD				
DEPARTMENT	Quality control department				
USER	Abel				
PRODUCE NO.	R01467				
PRODUCE NAME	Computer shell				

NO.	SEL	Name	20°	60°	85°	Sort	20° diff.	60° diff.	85° diff.	Result
1	☐	S001	-	14.7	-	Sample	-	-	-	-
2	☐	S002	-	0.7	-	Sample	-	-	-	-
3	☐	S003	-	99.6	-	Sample	-	-	-	-
4	☐	S004	-	99.6	-	Sample	-	-	-	-
5	☐	S005	-	100.7	-	Sample	-	-	-	-
6	☐	S006	-	15.2	-	Sample	-	-	-	-
7	☐	S007	-	0.5	-	Sample	-	-	-	-
8	☐	S008	-	100.7	-	Sample	-	-	-	-

Tolerance: 2.0

☐ Select All

Export Selected Information

Delete Selected Information

Print Report

Data Upload To PC

Measurement

Single measure: [Measure]

Aotu measure:
Time interval(s): 1 [Start]

● 20°
● 60°
● 85°

The current communication port is COM7, the communication port was opened! Instrument model: WG60G

☎ +982165565901
☎ +982144584619
✆ +989034119385