

The instrument must be operated continuously for more than 30 minutes at the specified temperature to ensure the following parameters.

Model		SP3051	SP3101	SP6101	SP6053	SP3103	SP6103
Rated Output (0°C-40°C)	Voltage	0 - 30V	0 - 30V	0 -60V	0 -60V	0 - 30V	0 -60V
	Current	5A	10A	10A	5A	10A	10A
	Output Power	150W	150W	150W	300W	300W	300W
Load Regulation	Voltage	≤30mV					
	Current	≤30mA					
Power Regulation	Voltage	≤20mV					
	Current	≤20mA					
Setting Resolution	Voltage	10mV					
	Current	1mA					
Readback Resolution	Voltage	10mV					
	Current	1mA					
Ripple/Noise (20Hz-20MHz)	Voltage (Vp-p)	≤30mVp-p		≤50mVp-p	≤30mVp-p		≤50mVp-p
	Voltage (rms)	≤3mVrms		≤5mVrms	≤3mVrms		≤5mVrms
	Current (rms)	≤30mA p-p					
Response Time	≤1.0ms						
Storage	5 groups of data						
Working Temperature	0-40°C						

The above parameters are subject to change without further notice, please refer to the website update.

+982165565901

+982144584619

+989034119385

پتروفرهان گستر جنوب

FGJ-NDT.IR

DIGINDT.IR