UVC LIGHT METER

Model: UVC-254A *ISO-9001, CE, IEC1010*







The Art of Measurement

يتروفرهان گستر جنوب

FGJ-NDT.IR

DIGINDT.IR

© +982165565901

© +982144584619

+989034119385

UVC LIGHT METER

Model: UVC-254A

Q Tehran, Tehransar



F	E	۸.	т	П	D	F	C
		٠.	ш	ч	к	c	-

- * Short wave 254 nm ultra-violet irradiance measurement.
- * Professional, high quality UVC meter.
- * Wide measurement range, 199.9 uW/cm 2 , 1.999 mW/cm 2 and 19.99 mW/cm 2 .
- * Microprocessor circuit assures maximum possible accuracy, provided special functions and features.
- Exclusive UVC sensor with correction filter.
- * Super large LCD display, easy readout.
- Heavy duty & compact housing case.
- * Records Maximum, Minimum with recall.
- * Data hold.
- * Auto power off saves battery life.
- * Operates from 9V battery.
- * RS232 PC serial interface.
- Zero adjustment by push button.

		GENERAL SPE	CIFICATIONS			
Circuit	Custom one-chip microprocessor LSI circuit.		Size	Main instrument: 180 x 72 x 32 mm		
Display	13 mm (0.5") large LCD display.			(7.1 x 2.8 x 1.3 inch).		
,	Dual function			Sensor probe:		
Spectral		Point 254 nm.		38 mm DIA. x 25 mm.		
response	Band pass	254 nm with narrow.	Accessories	Instruction manual		
range of		* Refer attached		UVC sensor probe 1 PC.		
sensor		spectrum.		Hard carrying case 1 PC.		
Measurement	199.9 uW/c	m^2 x 0.1 uW/cm^2	Optional	* Data Acquisition software,		
& ranges	10000 000 000 000 000 000 000 000 000 0	cm^2 x 0.001 mW/cm^2	accessories	SW-U801-WIN		
	The second secon	cm^2 x 0.01 mW/cm^2	decessories	* Excel data Acquisition software,		
Linearity	± 1%.		100	SW-E802		
Sensor		ve photo diode & UVC color		* RS232 cableUPCB-01		
	correction filter.			* USB cableUSB-01		
Memory	Records Ma	Records Maximum & Minimum		No. UV-C has been used for over 100		
Recall	reading with	reading with recall.		years in various applications to kill germ		
Zero Adj.	By push but	By push button.		on surfaces, in the air and in the water.		
Sample Time	Approx. 0.4	Approx. 0.4 sec.		UV-C has been proven to kill up to 99.9		
Power off	Manual off by push button, or auto			of germs. Many hospitals use UV-C to		
	shut off aft	shut off after 10 minutes approximately		sanitize. inks, adhesives and coatings.		
Over Range	Indication of ""		UVC sensor spectrum			
Indication				0.10		
Data output	RS 232/USB PC computer interface. * Connect the optional RS232 cable UPCB-02 will get the RS232 plug. * Connect the optional USB cable USB-01 will get the USB plug.		0.06 (WV) (0.06 0.04 0.04 0.05	0.08		
Operating Temperature and Humidity	Max. 80% F		visuods			
Power Supply		, MN1604 (PP3) or (Alkaline or Heavy duty		0.00		
Power Current	Approx. DC 5.3 mA.			200 240 280 320 Wavelength (nm)		
Weight	355 g/0.77	LB (included batteries)				

ELECTRICAL SPECIFICATIONS (23±5 ℃)						
Range	Resolution	Accuracy				
199.9 uW/cm^2	0.1 uW/cm^2	± (2 % FS + 2 dgt)				
1.999 mW/cm^2	0.001 mW/cm^2	FS : full scale				
19.99 mW/cm^2	0.01 mW/cm^2					

^{*} Appearance and specifications listed in this brochure are subject to change without notice.

1810-UVC254A

يتروفرهان كستر جنوب

FGJ-NDT.IR **DIGINDT.IR**



(Q) +982165565901



(c) +982144584619



+989034119385