



## Thermal hygrometer

### testo 608 - Continuous indoor climate monitoring

Continuous display of temperature, humidity and dewpoint

Max. and min. value display

Battery monitor

Additional advantages testo 608-H2:

LED alarm reports limit value violations

High accuracy  $\pm 2$  %RH



%RH

°C

°C td

The low-budget hygrometer testo 608-H1 continuously measures humidity, temperature and dewpoint. The large display is easily legible even at a distance, the suspension and standing fixtures allow flexible positioning on a table or the wall.

testo 608 has a max./min. value display and a battery monitor. Thanks to the long-term stable sensor, you can rely on correct measurement results even after years.

testo 608-H2, the precise alarm hygrometer, reliably reports humidity and temperature limit value violations, e.g. in garden centres, storerooms, cleanrooms, museums, laboratories etc.

+982165565901

+982144584619

+989034119385

پتروفرهان گستر جنوب

FGJ-NDT.IR

DIGINDT.IR

## Technical data / Accessories

### testo 608-H1

testo 608-H1 hygrometer, humidity/dewpoint/temperature measuring instrument with battery



Part no. 0560 6081

### testo 608-H2

testo 608-H2 hygrometer, humidity/dewpoint/temperature measuring instrument, with LED alarm, battery and calibration protocol



Part no. 0560 6082

#### General technical data

Measuring rate	18 s
Storage temp.	-40 to +70 °C
Battery type	9V block battery
Battery life	ca. 1 Jahr
Weight	168 g
Dimensions	111 x 90 x 40 mm
Material/Housing	ABS
Display	LCD, 2 lines
Warranty	2 years

#### Technical data

Sensor types	testo 608-H1		testo 608-H2	
	NTC	Testo humid. sensor, cap.	NTC	Testo humid. sensor, cap.
Meas. range	0 to +50 °C -20 to +50 °C td	+10 to +95 %RH	-10 to +70 °C -40 to +70 °C td	+2 to +98 %RH
Accuracy ±1 digit	±0.5 °C (at +25 °C)	±3 %RH (+10 to +95 %RH)	±0.5 °C (at +25 °C)	±2 %RH (+2 to +98 %RH)
Resolution	0.1 °C	0.1 %RH	0.1 °C	0.1 %RH
Oper. temp.	0 to +50 °C		-10 to +70 °C	

#### Accessories

Part no.

#### Accessories for measuring instrument

ISO calibration certificate humidity, Calibration points 11.3 %RH and 75.3 %RH at +25°C

0520 0006

+982165565901

+982144584619

+989034119385

پتروفرهان گستر جنوب

FGJ-NDT.IR

DIGINDT.IR